## Govt.of NCT of Delhi, Health & Family Welfare Deptt. 9th Level, 'A' Wing, Delhi Secretariat IP Estate New Delhi No. F.342/C-283/2012/H&FW | Ph Secylow 145-189 Dated: 4/2/2013

## **ORDER**

Whereas a Committee of officers was constituted by the Government of NCT of Delhi on 4/12/2012 to inquire into the facts and circumstances concerning the death of four patients admitted in the intensive care unit (ICU) of Sushrut Trauma Centre, New Delhi on 4/12/2012, due to alleged disruption/failure of supply of medical oxygen;

And whereas the said Committee submitted a detailed report after going into the facts and circumstances leading to the deaths of these four patients. The said Committee in its report at para 6(i) has held M/s P.E.S. Installation Pvt. Ltd accountable and responsible for not following standard procedures and failing in its duty to ensure safe supply of oxygen from manifold room to ICU, etc, and deploying untrained persons, in insufficient strength, to operate the manifold system at the relevant time;

And whereas an FIR in this regard vide no. 201/12 dated 4/12/2012 police station Civil Lines has also been lodged against the said firm under sections 287/304-A/304/34 of the Indian Penal Code;

Now, therefore, considering the facts and circumstances of this case in their totality, the Government of NCT of Delhi has taken a decision to hereby blacklist the said firm M/s P.E.S. Installation Pvt. Ltd and further debar it from participating in any tender process/ contract agreement with any of the Govt. Hospitals/Health Institutions of GNCTD, in respect of medical oxygen supply, for a period of two years from the date of this order.

Sd/-(S.C.L. Das) Secretary (H&FW)

To:

All Directors/Medical Supdt. of Hospitals/Health Institutions of GNCTD.

Copy for information to: M/s PES Installations Pvt. Ltd. 13 Km milestone Opp. Maharaja Agarsen Hospital Rohtak Road, Punjabi Bagh, New Delhi-110056

Deputy Secretary (H&FW)